




JWH-022

The Drug Enforcement Administration's Special Testing and Research Laboratory generated this monograph using structurally confirmed reference material.



## 1. GENERAL INFORMATION

**IUPAC Name:** naphthalen-1-yl[1-(pent-4-en-1-yl)-1H-indol-3-yl]methanone

**CAS#:** 209414-16-4

**Synonyms:** N/A

**Source:** DEA Reference Material Collection

**Appearance:** Light yellow powder

**UV<sub>max</sub>(nm):** Not Determined

## 2. CHEMICAL AND PHYSICAL DATA

### 2.1 CHEMICAL DATA

Form	Chemical Formula	Molecular Weight	Melting Point (°C)
Base	C <sub>24</sub> H <sub>21</sub> NO	339	56.4



## JWH-022

The Drug Enforcement Administration's Special Testing and Research Laboratory generated this monograph using structurally confirmed reference material.




### 3. QUALITATIVE DATA

#### 3.1 NUCLEAR MAGNETIC RESONANCE

**Sample Preparation:** Dilute analyte to ~10 mg/mL in DMSO-d<sub>6</sub> containing TMS for 0 ppm reference and dimethylfumarate as quantitative internal standard.

**Instrument:** 400 MHz NMR spectrometer  
**Parameters:** Spectral width: at least containing -3 ppm through 13 ppm  
Pulse angle: 90°  
Delay between pulses: 45 seconds  
<sup>1</sup>H NMR: JWH-022 Lot # ALB227-8; DMSO-d<sub>6</sub>; 400MHz






The Drug Enforcement Administration's Special Testing and Research Laboratory generated this monograph using structurally confirmed reference material.

### 3.2 GAS CHROMATOGRAPHY/MASS SPECTROMETRY

*Sample Preparation:* Dilute analyte ~4 mg/mL in methanol.

**Instrument:** Agilent gas chromatograph operated in split mode with MS detector  
**Column:** DB-1 MS (or equivalent); 30m x 0.25 mm x 0.25  $\mu$ m  
**Carrier Gas:** Helium at 1 mL/min  
**Temperatures:** Injector: 280°C  
MSD transfer line: 280°C  
MS Source: 230°C  
MS Quad: 150°C  
Oven program:  
1) 100°C initial temperature for 1.0 min  
2) Ramp to 300°C at 12 °C/min  
3) Hold final temperature for 9.0 min  
**Injection Parameters:** Split Ratio = 20:1, 1  $\mu$ L injected  
**MS Parameters:** Mass scan range: 30-550 amu  
Threshold: 100  
Tune file: stune.u  
Acquisition mode: scan  
**Retention Time:** 20.370 min

EI Mass Spectrum: JWH-022 Lot # ALB227-8







### 3.3 INFRARED SPECTROSCOPY (FTIR)

**Instrument:** FTIR with diamond ATR attachment (3 bounce)

**Scan Parameters:** Number of scans: 32  
Number of background scans: 32  
Resolution: 4 cm<sup>-1</sup>  
Sample gain: 8  
Aperture: 150

FTIR ATR (Diamond, 3 Bounce): JWH-022 Lot # ALB227-8





## JWH-022

*The Drug Enforcement Administration's Special Testing and Research Laboratory generated this monograph using structurally confirmed reference material.*



### **4. ADDITIONAL RESOURCES**

[Forendex](#)

[Wikipedia](#)